

## Element HT5

Paramètres	Unités	Chien	Chat	Lapin	Furet	Cheval	Vache	Cochon	Chèvre
<b>Globules Blancs</b>	10 <sup>3</sup> /μL	6.00 – 17.00	5.50 – 19.50	3.00 – 13.50	2.50 – 15.80	5.00 – 12.00	4.60 – 15.80	10.20 – 30.00	5.80 – 25.00
<b>Neu</b>	10 <sup>3</sup> /μL	3.62 – 12.30	3.12 – 12.58	0.50 – 6.60	0.60 – 10.80	2.18 – 6.96	0.6 – 4.9	2.80 – 16.10	2.12 – 10.10
<b>Lym</b>	10 <sup>3</sup> /μL	0.83 – 4.91	0.73 – 7.86	1.00 – 6.80	0.58 – 10.60	1.32 – 5.86	2.5 – 11.8	4.80 – 16.20	3.12 – 22.10
<b>Mon</b>	10 <sup>3</sup> /μL	0.14 – 1.97	0.07 – 1.36	0.08 – 1.51	0.00 – 1.12	0.05 – 0.92	0.00 – 1.02	0.20 – 2.25	0.00 – 1.42
<b>Eos</b>	10 <sup>3</sup> /μL	0.04 – 1.62	0.06 – 1.93	0.00 – 0.51	0.00 – 0.88	0.01 – 1.00	0.00 – 1.30	0.00 – 1.80	0.00 – 1.32
<b>Bas</b>	10 <sup>3</sup> /μL	0.00 – 0.12	0.00 – 0.12	0.00 – 0.76	0.00 – 0.20	0.00 – 0.12	0.00 – 0.35	0.00 – 0.46	0.00 – 0.35
<b>Neu</b>	%	52.0 – 81.0	38.0 – 80.0	14.0 – 62.0	16.5 – 81.3	38.0 – 70.0	7.1 – 38.2	20.0 – 63.0	13.0 – 58.0
<b>Lym</b>	%	12.0 – 33.0	12.0 – 45.0	25.0 – 82.0	15.3 – 72.8	25.0 – 62.0	52.3 – 85.6	30.0 – 70.0	35.0 – 83.0
<b>Mon</b>	%	2.0 – 13.0	1.0 – 8.0	2.0 – 15.0	0.0 – 8.8	1.0 – 8.0	0.0 – 9.5	1.0 – 12.0	0.0 – 11.0
<b>Eos</b>	%	0.5 – 10.0	1.0 – 11.0	0.0 – 6.0	0.0 – 6.9	0.1 – 8.0	0.0 – 9.6	0.0 – 7.0	0.0 – 8.0
<b>Bas</b>	%	0.0 – 1.3	0.0 – 1.2	0.0 – 8.0	0.0 – 1.8	0.0 – 1.2	0.0 – 1.9	0.0 – 2.0	0.0 – 2.5
<b>Globules Rouges</b>	10 <sup>6</sup> /μL	5.10 – 8.50	4.60 – 10.20	3.40 – 6.50	6.60 – 12.18	5.3 – 10.50	5.0 – 10.1	5.50 – 9.00	10.00 – 21.00
<b>Hémoglobine</b>	g/dl	11.0 – 19.0	8.5 – 15.3	8.0 – 14.0	12.5 – 18.2	10.0 – 17.0	8.0 – 14.2	10.0 – 16.0	6.2 – 13.5
<b>Hématocite</b>	%	33.0 – 56.0	26.0 – 47.0	25.0 – 42.0	37.5 – 59.0	30.0 – 49.0	23.0 – 42.5	33.0 – 52.0	19.0 – 36.0
<b>VGM</b>	fL	60.0 – 76.0	38.0 – 54.0	60.0 – 80.0	43.6 – 61.2	37.0 – 59.0	37.0 – 55.0	51.0 – 73.0	13.0 – 23.0
<b>TCMH</b>	pg	20.0 – 27.0	11.8 – 18.0	19.0 – 25.0	14.5 – 20.5	13.5 – 19.5	12.5 – 18.2	14.0 – 22.0	4.2 – 7.8
<b>CCMH</b>	g/dL	30.0 – 38.0	29.0 – 36.0	30.0 – 36.0	29.0 – 37.0	30.0 – 38.0	31.0 – 37.0	30.0 – 36.0	30.0 – 38.0
<b>IDR-CV</b>	%	12.5 – 17.2	16.0 – 23.0	10.5 – 22.0	13.1 – 20.3	19.5 – 25.0	17.5 – 26.5	17.0 – 29.5	18.0 – 30.0
<b>PLT</b>	10 <sup>3</sup> /μL	117 – 490	100 – 518	100 – 1250	238 – 910	90 – 360	100 – 720	200 – 1000	–
<b>VMP</b>	fL	8.0 – 14.1	9.9 – 16.3	4.0 – 7.8	5.8 – 11.6	5.5 – 8.0	4.8 – 7.6	7.2 – 13.1	–

## Element HT5

Paramètres	Unités	Mouton	Lama	Souris	Rat	Singe	Panda géant	Panda rouge
<b>Globules Blancs</b>	10 <sup>3</sup> /μL	5.10 – 15.80	9.10 – 29.20	0.80 – 10.60	1.90 – 16.80	8.00 – 27.00	5.80 – 15.00	5.10 – 15.00
<b>Neu</b>	10 <sup>3</sup> /μL	1.32 – 8.96	3.80 – 23.60	0.23 – 3.60	0.35 – 6.30	2.00 – 12.10	3.00 – 13.10	2.50 – 8.20
<b>Lym</b>	10 <sup>3</sup> /μL	2.01 – 7.80	1.52 – 7.60	0.60 – 8.90	0.91 – 12.20	3.80 – 14.20	0.90 – 3.80	1.00 – 8.00
<b>Mon</b>	10 <sup>3</sup> /μL	0.00 – 1.52	0.00 – 1.80	0.04 – 1.40	0.08 – 2.30	0.19 – 2.60	0.32 – 1.60	0.20 – 1.40
<b>Eos</b>	10 <sup>3</sup> /μL	0.00 – 1.08	0.00 – 3.80	0.00 – 0.51	0.00 – 1.01	0.00 – 1.14	0.00 – 1.40	0.00 – 0.82
<b>Bas</b>	10 <sup>3</sup> /μL	0.00 – 0.17	0.00 – 0.23	0.00 – 0.12	0.00 – 0.20	0.00 – 0.29	0.00 – 0.10	0.00 – 0.10
<b>Neu</b>	%	21.5 – 68.0	38.5 – 85.6	6.5 – 50.0	7.3 – 50.0	20.0 – 50.0	50.0 – 80.0	30.0 – 70.0
<b>Lym</b>	%	28.0 – 64.5	11.2 – 42.5	40.0 – 92.0	40.0 – 88.9	40.0 – 70.0	12.0 – 36.0	20.0 – 65.0
<b>Mon</b>	%	0.0 – 14.3	0.0 – 10.3	0.9 – 18.0	2.0 – 18.0	2.0 – 12.0	2.0 – 14.0	2.0 – 12.0
<b>Eos</b>	%	0.0 – 8.0	0.0 – 18.0	0.0 – 7.5	0.0 – 7.0	0.0 – 7.0	0.0 – 9.0	0.0 – 8.0
<b>Bas</b>	%	0.0 – 1.5	0.0 – 1.5	0.0 – 1.5	0.0 – 1.5	0.0 – 1.5	0.0 – 1.0	0.0 – 1.0
<b>Globules Rouges</b>	10 <sup>6</sup> /μL	5.50 – 14.20	8.12 – 15.62	6.50 – 11.50	5.00 – 9.8	4.50 – 6.80	5.00 – 8.00	6.50 – 9.50
<b>Hémoglobine</b>	g/dl	6.3 – 13.2	8.0 – 15.0	11.0 – 16.5	11.0 – 17.0	10.0 – 16.0	10.0 – 15.5	10.0 – 15.5
<b>Hématocite</b>	%	20.0 – 39.0	20.0 – 36.0	35.0 – 55.0	32.0 – 53.0	33.0 – 48.0	30.0 – 45.0	30.0 – 45.0
<b>VGM</b>	fL	25.0 – 41.0	20.0 – 29.0	41.0 – 55.0	50.0 – 67.0	68.0 – 82.0	50.0 – 62.0	41.0 – 52.0
<b>TCMH</b>	pg	8.0 – 12.3	8.5 – 13.2	13.0 – 18.0	16.0 – 23.0	19.0 – 27.0	18.0 – 23.0	13.5 – 17.0
<b>CCMH</b>	g/dL	29.0 – 36.0	34.0 – 47.0	30.0 – 36.0	31.0 – 37.0	30.0 – 36.0	32.0 – 37.0	31.0 – 36.0
<b>IDR-CV</b>	%	16.5 – 26.2	18.2 – 25.5	12.0 – 19.0	11.0 – 16.0	11.0 – 16.2	14.0 – 19.0	13.0 – 18.0
<b>PLT</b>	10 <sup>3</sup> /μL	100 – 800	200 – 700	400 – 1600	250 – 1500	200 – 800	580 – 1350	200 – 700
<b>VMP</b>	fL	3.5 – 6.0	2.3 – 4.2	4.0 – 6.2	4.8 – 7.5	7.2 – 12.0	3.8 – 6.5	6.5 – 10.0